



U Unicare Clear Vinyl (GS006*-A)

clear \cdot medical grade \cdot medium weight \cdot vinyl examination gloves

Physical Specifications:

Material: Vinyl (latex free) Colour: Clear Length: 240 mm Weight: 4.5 g Thickness: 0.06 mm Finish: Powder free Fit: Ambidextrous Sterile: No

Order Codes and Size Information:

GS0062-A | Small GS0063-A | Medium GS0064-A | Large GS0065-A | Extra large

Packaging:

Each dispenser box contains 100 gloves Each case contains 10 dispenser boxes

Classification:



For more about the Unicare range, please click on the image



UNIGLOVES UK

unialoves.co.uk

Registered in England No: 4010200 VAT registration 791817787

+44 (0)800 049 6602

enquiries@unigloves.co.uk

3 Ambley Green, Gillingham Business Park, Gillingham, ME8 ONJ, UK

Features

- AQL 1.5 and EN455 parts 1-4
- Free from latex and accelerators
- Free from powder
- Tested to EN1186
- Ambidextrous
- Manufactured using high-quality vinyl
- Beaded cuff
- Medium-weight construction

Benefits

- Meets medical standards including pinholes
- Suitable for those with latex allergies (Type I) sensitisation) and Type IV hypersensitivity
- Reduces risks of latex allergies (Type I sensitisation) and minimises particulate contamination
- Suitable for contact with non-fatty food types
- Ambidextrous fit means you don't have to split a pair to replace a damaged glove
- High-quality vinyl combines low cost with reduced risk of protein allergens
- Beaded cuff provides extra strength during donning and helps prevent liquid roll back
- Medium-weight construction provides a balance of strength and sensitivity

C Sectors

- Care home work
- Cleaning and janitorial
- Dentistry
- Dental hygiene
- Emergency services
- Food processing
- General assembly
- Healthcare
- Hospital and clinical duties
- lanitorial
- Local Authority work
- Painting and decorating
- Printing

UNIGLOVES GERMANY Arzt- & Klinikbedarf Handelsgesellschaft mbH

Camp-Spich-Straße 71 D-53842 Troisdorf-Spich, +49 (0) 2241-9323-0 unigloves@unigloves.de unialoves.de



Issue No. 6

Prepared at Gillingham on 08/03/22



